



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/811,565	03/29/2004	Salvador Marquez	ECV-5783	1380

30452 7590 04/19/2007  
EDWARDS LIFESCIENCES CORPORATION  
LEGAL DEPARTMENT  
ONE EDWARDS WAY  
IRVINE, CA 92614

EXAMINER
----------

PRONE, CHRISTOPHER D

ART UNIT	PAPER NUMBER
----------	--------------

3738

SHORTENED STATUTORY PERIOD OF RESPONSE	MAIL DATE	DELIVERY MODE
3 MONTHS	04/19/2007	PAPER

**Please find below and/or attached an Office communication concerning this application or proceeding.**

If NO period for reply is specified above, the maximum statutory period will apply and will expire 6 MONTHS from the mailing date of this communication.

ED

<b>Office Action Summary</b>	<b>Application No.</b> 10/811,565	<b>Applicant(s)</b> MARQUEZ, SALVADOR	
	<b>Examiner</b> Christopher D. Prone	<b>Art Unit</b> 3738	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

#### Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

#### Status

- 1) ☒ Responsive to communication(s) filed on 05 February 2007.
- 2a) ☐ This action is FINAL. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11; 453 O.G. 213.

#### Disposition of Claims

- 4) ☒ Claim(s) 1-23 is/are pending in the application.
- 4a) Of the above claim(s) 19-23 is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1-18 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

#### Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

#### Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some \* c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
  - ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

#### Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☒ Information Disclosure Statement(s) (PTO/SB/08)  
Paper No(s)/Mail Date 6/18/04 6/18/04 10/14/05

- 4) ☐ Interview Summary (PTO-413)  
Paper No(s)/Mail Date. \_\_\_\_\_
- 5) ☐ Notice of Informal Patent Application
- 6) ☐ Other: \_\_\_\_\_

## **DETAILED ACTION**

### ***Election/Restrictions***

Applicant's election of Invention 1 claims 1-18 in the reply filed on 2/5/07 is acknowledged. Because applicant did not distinctly and specifically point out the supposed errors in the restriction requirement, the election has been treated as an election without traverse (MPEP § 818.03(a)).

### ***Specification***

The abstract of the disclosure is objected to because it has too many words. The current abstract comprises 173 words, but is limited to 150. Correction is required. See MPEP § 608.01(b).

### ***Claim Rejections - 35 USC § 102***

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) The invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1, 5, 6, 9, and 10 are rejected under 35 U.S.C. 102(b) as being anticipated by USPN 6,736,845 B2 Marquez et al.

Marquez discloses the same invention comprising a support frame for a flexible leaflet prosthetic heart valve made of nitinol comprising a plurality of

Art Unit: 3738

cusps 74, a plurality of enlarged ear shaped commissures 80 further comprising a fragile bridge 90, and a fabric covering

***Claim Rejections - 35 USC § 103***

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 2-4, 8, and 11-18 are rejected under 35 U.S.C. 103 as being unpatentable over Marquez in view of Love.

Marquez discloses the invention substantially as claimed being described above. However, Marquez does not disclose that the frame and bridges be a continuous homogeneous material.

The examiner is taking official notice that it is obvious and well known in the art of product manufacturing to minimize the amount of parts and materials during production in order to reduce costs and simplify the production process. Therefore it would be obvious to produce the implant of the Marquez patent having a continuous homogenous frame in order to simplify the production process. In support for a continuous homogenous support frame is Love, which further discloses a plurality of commissures each comprising weak notched out sections shown best in figure 16. Love teaches the use of a continuous

Art Unit: 3738

homogeneous notched frame Figure 16 in the same field of endeavor for the purpose of providing a flexible yet rigid support frame for a heart valve.

It would have been obvious to one having ordinary skill in the art at the time the invention was made to combine the teaching of a continuous homogeneous support frame as supported by Love with the implant of Marquez in order to simplify production and reduce costs.

### ***Conclusion***

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Christopher D. Prone whose telephone number is (571) 272-6085. The examiner can normally be reached on Monday Through Fri 8:30 to 5:00.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Corrine McDermott can be reached on (571) 272-4754. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3738

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

  
CDP

Christopher D Prone  
Examiner  
Art Unit 3738

  
**CORRINE McDERMOTT**  
**SUPERVISORY PATENT EXAMINER**  
**TECHNOLOGY CENTER 3700**